



CARBON DIOXIDE REGULATORS

- Available with any international high pressure outlet / probe fitting
- Luminous gauge face for night use
- Service free over the products life time

Outlet Type:

7-Year warranty

Meditech offers a wide range of high-pressure regulators for medical use and within this is our CO2 models. They are available for use with any outlet / inlet fitting needed to meet local requirements / specifications.

SPECIFICATION 200 bar Rated Inlet Pressure: 4 bar (400 kPa) Storage temperature range: -40°C to +60°C Operating temperature range: -20°C to +60°C Intended Life: 10 Years Service Interval: Service free over products life time Warranty Period: 7 Years Inlet Type: Pin Index
Others available upon request

BS Schrader

Others available upon request



B.N.O.S. Meditech Ltd.

9 Fifth Avenue, Bluebridge Industrial Estate, Halstead, Essex CO9 2SZ, England, UK.

Tel: +44 (0)1787 479475 • Fax: +44 (0)1787 477747

E-mail: sales@meditech.uk.com • www.meditech.uk.com

TO PURCHASE OR FOR FURTHER DETAILS: Contact us by email or telephone for more details, including specific part numbers, prices, etc.

B.N.O.S. Meditech Ltd.
9 Fifth Avenue, Bluebridge Industrial Estate,
Halstead, Essex CO9 2SZ, England, UK.
Tel: +44 (0)1787 479475 • Fax: +44 (0)1787 477747

E-mail: sales@meditech.uk.com • www.meditech.uk.com

Meditech Regulators are CE marked in accordance with the European Medical Devices Directive 93/42/EEC and UKCA marked accordance with the UK Medical Devices Regulations 2002. All items in this range are sold with a return to manufacturer warranty against manufacturing defects. These products are part of a series of medical gas and resuscitation products available from B.N.O.S. Meditech. If the product you require is not shown here, or if you require more information, please contact your distributor or contact us directly. Specifications may be subject to change without notice.

B.N.O.S. Meditech Ltd Terms and Conditions of sale apply, available on application or from our website.





Distributor:

Doc 946M Issue E 16/08/2024